



Internet of Things Enabled Smart Healthcare Monitoring System Using Wireless Sensors

Jovencio Saeedzada

Department of Electrical and Computer Engineering, Angkor Mekong Technical University, Cambodia

jovencio.saeedzada@amtu-kh.edu

Peer Review Information

Submission: 19 May 2021

Revision: 04 June 2021

Acceptance: 20 June 2021

Keywords

*Internet of Things, Smart
healthcare, Wireless
sensors, Remote patient
monitoring, Wearable
health monitoring.*

Abstract

Internet of Things (IoT)-enabled healthcare monitoring combines physiological sensors, low-power wireless communication, gateways and remote software services to support continuous observation outside conventional clinical settings. This methodology paper proposes a structured experimental framework for an IoT-enabled smart healthcare monitoring system using wireless sensors. The system acquires heart rate, peripheral oxygen saturation (SpO₂), body temperature, electrocardiogram (ECG) or activity data and, where appropriate, blood-pressure measurements through wearable or bedside sensor nodes. Sensor readings are transmitted through Bluetooth Low Energy (BLE), ZigBee or Wi-Fi to a gateway, processed locally for validation and threshold-based alerting, and forwarded to a secure cloud or hospital dashboard for longitudinal monitoring. Foundational IoT studies emphasize the convergence of wireless sensor networks, Internet connectivity and distributed computing, while healthcare-specific surveys demonstrate the importance of wearable sensing, communication standards, privacy and interoperable architectures. IoT-aware hospital architectures have shown that wireless sensors can support continuous collection of patient and environmental data and edge/fog gateways can reduce latency and communication overhead in healthcare IoT. The proposed methodology evaluates sensing accuracy, packet-delivery ratio, end-to-end latency, alert delay, data availability, wireless-node energy consumption and gateway resource utilization under normal, mobility and packet-loss conditions. A prototype can be implemented using microcontroller-based sensor nodes and an edge gateway, while a patient-monitoring dataset or calibrated reference devices are used for validation. Because no patient dataset or hardware logs were supplied, the Results section provides illustrative expected patterns rather than fabricated clinical measurements. The framework is intended for engineering evaluation and not for diagnosis or replacement of certified medical equipment.

Introduction

Healthcare systems increasingly require continuous and remote observation of patients with chronic disease, elderly individuals and people recovering outside hospitals. Conventional periodic measurements may miss

short-duration abnormalities, while frequent hospital visits increase cost and inconvenience. Wireless wearable and environmental sensors provide a way to collect longitudinal

physiological information in daily life and transmit it to caregivers when needed.

The IoT extends wireless sensing by connecting sensor nodes, gateways, applications and healthcare services through interoperable networks. Gubbi et al. (2013) described IoT as a convergence of sensing, networking and distributed computation. In healthcare, Islam et al. (2015) showed that IoT architectures can integrate body sensors, communications, services and security functions, while Hassanaliieragh et al. (2015) demonstrated the potential of networked health sensors combined with cloud-based processing.

Literature Review

Early IoT-healthcare research established the architectural role of wireless sensors and connected services. Catarinucci et al. (2015) proposed an IoT-aware smart-hospital architecture integrating wireless sensor networks, RFID and mobile technologies to monitor physiological and environmental parameters. Their work illustrates the importance of a gateway layer that can connect constrained sensing devices to Internet services. Hassanaliieragh et al. (2015) similarly considered wearable and environmental sensing linked to cloud processing for health monitoring and management.

Standardization, interoperability and low-power operation remain important constraints. Baker et al. (2017) reviewed IoT technologies for smart healthcare and identified wearability, security, privacy, communication standards and energy consumption as major design issues. Majumder et al. (2017) reviewed wearable sensors for

remote health monitoring and emphasized continuous observation of vital signs in cost-effective systems. Wu et al. (2017) demonstrated an autonomous wireless body area network implementation for IoT-connected healthcare, reinforcing the need to jointly evaluate sensing quality and communication/energy performance.

Edge and fog computing were introduced to reduce dependence on distant cloud servers. Rahmani et al. (2018) proposed smart e-health gateways that execute local processing at the edge of healthcare IoT, while Farahani et al. (2018) presented a fog-driven e-health architecture spanning devices, fog resources and cloud services. These approaches are especially relevant for physiological alerts because an abnormal reading should not depend entirely on wide-area network availability before a warning can be generated.

Rodrigues et al. (2018) reviewed enabling technologies for the Internet of Health Things, including sensors, communications and ambient-assisted living. Selvaraj and Sundaravaradhan (2020) later identified accuracy, power consumption, data management, security and cloud integration as recurring challenges in IoT healthcare systems. Kadhim et al. (2020) reviewed patient health-status monitoring based on IoT and highlighted the role of wearables, wireless channels and remote diagnosis. More recent wearable-sensor research through 2021 has continued to emphasize multimodal sensing, wireless connectivity and practical constraints such as energy, comfort and reliability (Mamdiwar et al., 2021).

Table 1. Selected Literature on IoT and Wireless-Sensor Healthcare Monitoring

Study	Focus	Contribution to proposed methodology
Islam et al. (2015)	Comprehensive IoT-healthcare survey	Defines architecture, communication, applications and security requirements.
Catarinucci et al. (2015)	IoT-aware smart hospital	Supports wireless physiological/environmental sensing and gateway integration.
Baker et al. (2017)	Smart healthcare technologies	Highlights standards, wearability, security and low-power design.
Majumder et al. (2017); Wu et al. (2017)	Wearable sensors / WBAN	Supports continuous vital-sign sensing and low-power wireless links.
Rahmani et al. (2018); Farahani et al. (2018)	Edge/fog healthcare IoT	Motivates local filtering and low-latency alert generation.
Selvaraj & Sundaravaradhan (2020); Mamdiwar et al. (2021)	IoT-healthcare reviews	Identifies accuracy, energy, reliability and usability as key evaluation dimensions.

Many studies separately discuss wearable sensors, IoT architecture or cloud services. A

practical monitoring system, however, must demonstrate that physiological values are

sufficiently accurate, wireless transmission is reliable, alerts arrive quickly and battery-powered nodes remain energy efficient. The proposed study therefore evaluates the complete sensing-to-alert chain rather than only the sensor or application layer.

The proposed study has the following objectives:

design an IoT-enabled smart healthcare monitoring architecture using wearable or bedside wireless sensors;
measure physiological parameters including heart rate, SpO₂, body temperature and optional ECG/activity or blood-pressure data;
evaluate sensing accuracy against calibrated reference instruments or labelled validation data;
measure packet-delivery ratio, end-to-end latency, alert delay and data availability under different wireless conditions;
estimate sensor-node energy consumption and gateway resource utilization; and
compare cloud-only alert processing with gateway/edge-assisted local alerting.

The research questions are: RQ1: Can the proposed wireless sensor system acquire and transmit physiological values with acceptable

accuracy and availability? RQ2: How do distance, packet loss and mobility affect latency and packet-delivery ratio? RQ3: Does local gateway processing reduce alert delay when Internet connectivity is degraded? RQ4: Which sensing/communication configuration provides the best trade-off between monitoring quality and energy consumption?

Methodology

1. Research Design

A prototype-based quantitative experimental design is proposed. The study evaluates both biomedical-sensing performance and network/computing performance. Healthy-volunteer or bench-test measurements may be used only after appropriate institutional/ethical approval; alternatively, sensor signals can be validated using simulators, publicly available physiological recordings or calibrated reference instruments. The system is not intended to make clinical diagnoses. Each technical condition should be repeated multiple times so that mean, standard deviation and confidence intervals can be reported.




Figure 1. IoT-Enabled Smart Healthcare Monitoring Architecture Using Wireless Sensors

2. Hardware and Communication Architecture

The sensing layer consists of a low-power microcontroller and compatible sensors. A photoplethysmography module can provide heart rate and SpO₂; a digital temperature sensor can measure body or skin temperature; ECG electrodes or an inertial sensor can be added for cardiac/activity monitoring. The node transmits

timestamped readings over BLE or ZigBee to a smartphone, Raspberry Pi-class gateway or equivalent embedded platform. Wi-Fi/Ethernet connects the gateway to the cloud. This layered architecture follows the sensor-gateway-service pattern described across smart-healthcare literature.

3. Variables and Measurement Plan

Table 2. Sensors, Variables and Validation Measures

Parameter component /	Measurement source	Validation / KPI
Heart rate	PPG / ECG-derived sensor	Mean absolute error (bpm); correlation with reference.
SpO ₂	Pulse-oximetry sensor	Absolute error (%); valid-reading rate.
Temperature	Digital temperature sensor	Absolute error (°C); stability.

ECG / activity (optional)	ECG module / IMU	Signal availability, event/activity agreement.
Wireless link	BLE / ZigBee / Wi-Fi	PDR, latency, RSSI, retransmissions.
Alert mechanism	Gateway + cloud	Threshold-detection time and end-to-end alert delay.
Power	Current/voltage logging	Average power and estimated battery life.

4. Data Acquisition and Preprocessing

Each sensor reading is timestamped at the node and associated with a device identifier and signal-quality flag. Sampling frequency is selected according to the parameter: low-frequency vital signs such as temperature require fewer samples, whereas ECG requires substantially higher sampling. The gateway removes malformed packets, checks physiological plausibility, applies light smoothing or median filtering where appropriate, and stores both raw and processed values. Missing values are not silently replaced; missingness and packet loss are separately

quantified. Threshold-based local alerts are configured for engineering testing using predefined safe test ranges or simulated abnormal values. Clinical threshold selection should be performed only by qualified healthcare professionals in an approved study. The gateway generates a local notification immediately and then synchronizes the event with the cloud. This architecture reflects the edge/fog rationale described by Rahmani et al. (2018) and Farahani et al. (2018).

5. Experimental Scenarios

Table 3. Proposed Experimental Scenarios

Scenario	Condition	Purpose
S1 Baseline	Sensor 1-2 m from gateway; stable link	Establish accuracy, latency and PDR baseline.
S2 Distance	Increase separation / introduce indoor obstruction	Measure RSSI, PDR and retransmission effects.
S3 Mobility	Wearable node moves between rooms / locations	Assess continuity and temporary disconnection.
S4 Congestion	Multiple sensor nodes transmit simultaneously	Evaluate gateway/network scalability.
S5 Internet impairment	Add WAN delay or disconnect cloud link	Compare cloud-only vs local edge alerting.
S6 Battery test	Continuous monitoring over extended period	Estimate energy consumption and battery lifetime.

Step-by-Step Experimental Procedure

Step 1 - Calibrate every sensor using a reference instrument, simulator or known test input.

Step 2 - Configure node identifiers, sampling rates, wireless protocol, timestamps and secure gateway registration.

Step 3 - Collect synchronized sensor and reference measurements under the baseline condition.

Step 4 - Calculate sensor error, valid-reading percentage and short-term stability for each physiological parameter.

Step 5 - Run wireless-distance, mobility and multi-node scenarios while recording sent/received packet counts, RSSI and timestamps.

Step 6 - Calculate packet-delivery ratio, end-to-end latency, jitter and periods of unavailable data.

Step 7 - Inject simulated abnormal values and compare local gateway alert delay with cloud-only alert delay.

Step 8 - Record current consumption in sensing, transmission, idle and alert states and estimate battery life.

Step 9 - Repeat each scenario at least 10 times or for a sufficiently long monitoring interval to obtain stable estimates.

Step 10 - Summarize results using mean, standard deviation, 95% confidence intervals and appropriate statistical comparisons.

6. Performance Metrics and Statistical Analysis

Sensor accuracy is reported using mean absolute error (MAE), root mean square error where useful, correlation and valid-reading rate. Network performance is reported as packet-delivery ratio (received packets/sent packets $\times$ 100), end-to-end delay, jitter and outage

duration. Alert performance is measured from abnormal-value detection to local and remote notification. Energy performance is represented by average power and estimated operating hours from a specified battery capacity. Paired tests can compare edge and cloud alert latency; ANOVA or a non-parametric equivalent can compare multiple distance/load conditions. Statistical significance should be accompanied by effect sizes and confidence intervals.

Results

No patient dataset, calibrated-device measurements or wireless packet logs were supplied with this request. Therefore, the numerical values below are illustrative examples that demonstrate the expected reporting structure and must not be represented as actual clinical or experimental findings.

Table 4. Illustrative System Performance

Metric	Baseline	Challenging condition	Expected interpretation
Heart-rate MAE	1.8 bpm	3.2 bpm during motion	Motion artifact increases sensing error.
SpO ₂ MAE	1.1%	2.0% during motion	Signal quality affects optical measurement.
Packet-delivery ratio	99.4%	96.2% with obstruction/mobility	Wireless degradation produces missing samples.
Median end-to-end latency	48 ms	126 ms under congestion	Load and retransmission increase delay.
Local edge alert delay	85 ms	110 ms	Remains available without Internet.
Cloud alert delay	310 ms	>2 s / unavailable during WAN fault	WAN conditions dominate remote alert timing.
Estimated node lifetime	3.8 days	2.9 days at high transmit rate	Communication frequency strongly affects energy use.

The first expected result is that sensor accuracy remains high under stationary conditions but degrades during motion or poor sensor contact. Wearable-health studies emphasize that practical monitoring must account for comfort, placement, signal quality and artifact rather than assuming laboratory accuracy in all conditions (Majumder et al., 2017; Mamdiwar et al., 2021). For this reason, the study reports valid-reading rate in addition to numerical error. The second expected result is a trade-off between wireless reliability and energy consumption. Increasing transmission frequency can reduce the time between measurements and alerts but raises radio duty cycle and battery drain. Baker et al. (2017), Wu et al. (2017), and Selvaraj and Sundaravaradhan (2020) identify low-power operation as a core challenge for connected healthcare. The most useful configuration is therefore not necessarily the fastest one; it should meet the application's maximum acceptable delay while preserving practical battery life. The third expected result is that gateway/edge alerting remains responsive when the cloud path is congested or unavailable. Rahmani et al. (2018) and Farahani et al. (2018) argue that healthcare gateways can perform useful local processing rather than operating only as protocol translators. In the proposed system, threshold checking and short local buffers allow

monitoring to continue through temporary Internet outages, after which stored data are synchronized to the cloud. Finally, the multi-node experiment is expected to reveal a point at which packet collisions, retransmissions or gateway processing load begin to reduce data availability. This scalability behaviour is important because a hospital ward or assisted-living environment may contain many simultaneously active sensors. The results should therefore present PDR and delay as functions of node count rather than only a single-device benchmark.

Discussion

The proposed methodology evaluates the complete IoT healthcare chain from sensing to notification. This is important because a system with accurate sensors can still fail if wireless transmission is unreliable, while a low-latency network cannot compensate for poor physiological measurements. The framework therefore combines biomedical signal error, network quality, local processing, power consumption and alert delivery. Earlier architectures established that connected healthcare requires integration among constrained sensor devices, gateways, Internet services and clinical interfaces (Islam et al., 2015; Catarinucci et al., 2015; Rodrigues et al., 2018).

The present design follows this layered model while adding comparative experiments for edge and cloud processing. It also recognizes security and privacy as design requirements; data should be encrypted in transit, devices authenticated, access logged and personally identifiable information minimized. These safeguards are engineering requirements but do not replace formal compliance review for a real clinical deployment.

Conclusion

This methodology paper presents a compact and reproducible framework for an Internet of Things enabled smart healthcare monitoring system using wireless sensors. The proposed architecture uses wearable or bedside sensor nodes to acquire physiological data, a low-power wireless link to a local gateway, edge logic for filtering and urgent alerts, and cloud/hospital services for long-term storage and remote visualization. Its main contribution is the simultaneous evaluation of sensing accuracy, wireless reliability, latency, data availability, alert response and energy consumption. The expected engineering outcome is that local gateway processing improves resilience and alert responsiveness, particularly when the Internet connection is slow or unavailable, while wireless-sensor optimization is required to balance data freshness and battery life. Continuous monitoring may improve visibility of physiological trends, but any real patient study must use validated hardware, clinician-defined thresholds, informed consent and appropriate ethical approval. The prototype should therefore be considered a research and monitoring platform rather than a diagnostic medical device. Future work can incorporate secure federated analytics, anomaly detection, 5G/LPWAN connectivity, adaptive sampling and clinical decision-support models. Longer trials should evaluate comfort, sensor drift, packet loss in realistic homes or wards and the reliability of emergency escalation. The methodology nevertheless provides a clear step-by-step basis for testing IoT healthcare monitoring systems using wireless sensors.

References

Gubbi, J., Buyya, R., Marusic, S., & Palaniswami, M. (2013). Internet of Things (IoT): A vision, architectural elements, and future directions. *Future Generation Computer Systems*, 29(7), 1645-1660.
<https://doi.org/10.1016/j.future.2013.01.010>

Islam, S. M. R., Kwak, D., Kabir, M. H., Hossain, M., & Kwak, K.-S. (2015). The Internet of Things for health care: A comprehensive survey. *IEEE Access*, 3, 678-708.
<https://doi.org/10.1109/ACCESS.2015.2437951>

Catarinucci, L., de Donno, D., Mainetti, L., Palano, L., Patrono, L., Stefanizzi, M. L., & Tarricone, L. (2015). An IoT-aware architecture for smart healthcare systems. *IEEE Internet of Things Journal*, 2(6), 515-526.
<https://doi.org/10.1109/JIOT.2015.2417684>

Hassanalieragh, M., Page, A., Soyata, T., Sharma, G., Aktas, M., Mateos, G., Kantarci, B., & Andreescu, S. (2015). Health monitoring and management using Internet-of-Things (IoT) sensing with cloud-based processing: Opportunities and challenges. *2015 IEEE International Conference on Services Computing*, 285-292.
<https://doi.org/10.1109/SCC.2015.47>

Baker, S. B., Xiang, W., & Atkinson, I. (2017). Internet of Things for smart healthcare: Technologies, challenges, and opportunities. *IEEE Access*, 5, 26521-26544.
<https://doi.org/10.1109/ACCESS.2017.2775180>

Majumder, S., Mondal, T., & Deen, M. J. (2017). Wearable sensors for remote health monitoring. *Sensors*, 17(1), 130.
<https://doi.org/10.3390/s17010130>

Wu, T., Wu, F., Redoute, J.-M., & Yuce, M. R. (2017). An autonomous wireless body area network implementation towards IoT connected healthcare applications. *IEEE Access*, 5, 11413-11422.
<https://doi.org/10.1109/ACCESS.2017.2716344>

Rahmani, A. M., Gia, T. N., Negash, B., Anzanpour, A., Azimi, I., Jiang, M., & Liljeberg, P. (2018). Exploiting smart e-Health gateways at the edge of healthcare Internet-of-Things: A fog computing approach. *Future Generation Computer Systems*, 78, 641-658.
<https://doi.org/10.1016/j.future.2017.02.014>

Farahani, B., Firouzi, F., Chang, V., Badaroglu, M., Constant, N., & Mankodiya, K. (2018). Towards fog-driven IoT eHealth: Promises and challenges of IoT in medicine and healthcare. *Future Generation Computer Systems*, 78, 659-676.
<https://doi.org/10.1016/j.future.2017.04.036>

Rodrigues, J. J. P. C., De Rezende Segundo, D. B., Junqueira, H. A., Sabino, M. H., Prince, R. M. I., Al-Muhtadi, J., & de Albuquerque, V. H. C. (2018). Enabling technologies for the Internet of Health

Things. IEEE Access, 6, 13129-13141.
<https://doi.org/10.1109/ACCESS.2017.2789329>

Selvaraj, S., & Sundaravaradhan, S. (2020). Challenges and opportunities in IoT healthcare systems: A systematic review. SN Applied Sciences, 2, 139.
<https://doi.org/10.1007/s42452-019-1925-y>

Kadhim, K. T., Alsahlany, A. M., Wadi, S. M., & Kadhum, H. T. (2020). An overview of patient's

health status monitoring system based on Internet of Things (IoT). Wireless Personal Communications, 114, 2235-2262.
<https://doi.org/10.1007/s11277-020-07474-0>

Mamdiwar, S. D., R, A., Shakruwala, Z., Chadha, U., Srinivasan, K., & Chang, C.-Y. (2021). Recent advances on IoT-assisted wearable sensor systems for healthcare monitoring. Biosensors, 11(10), 372.
<https://doi.org/10.3390/bios11100372>